# **DRAFT Medical Coverage Policy** | Corneal

Topography/Computer-Assisted Corneal Topography/Photokeratoscopy



**EFFECTIVE DATE:** 04 | 01 | 2025

**POLICY LAST REVIEWED:** 12 | 18 | 2024

#### **OVERVIEW**

Computer-assisted corneal topography (also called photokeratoscopy or videokeratography) provides a quantitative measure of corneal curvature. Measurement of corneal topography is being evaluated to aid the diagnosis of and follow-up for corneal disorders such as keratoconus, difficult contact lens fits, and pre- and postoperative assessment of the cornea, most commonly after refractive surgery.

### **MEDICAL CRITERIA**

Not applicable

#### **PRIOR AUTHORIZATION**

Not applicable

### **POLICY STATEMENT**

## Medicare Advantage Plans and Commercial Products

Computerized corneal topography is considered medically necessary under any of the following conditions, and when filed with one of the covered ICD-10 codes listed under the Coding section:

- Pre-operative evaluation of irregular astigmatism for intraocular lens power determination with cataract surgery
- Monocular diplopia
- Diagnosis of early keratoconus
- Post-surgical or post-traumatic astigmatism, measuring at a minimum of 3.5 diopters
- Suspected irregular astigmatism based on retinoscopic streak or conventional keratometry
- Post-penetrating keratoplasty surgery
- Post-surgical or post-traumatic irregular astigmatism
- Certain corneal dystrophies
- Complications of transplanted cornea
- Post-traumatic corneal scarring
- Pterygium and/or corneal ectasia that cause visual impairment

Corneal topography considered medically necessary for pre-operative cataract assessment only if the individual has irregular astigmatism. Its use for this purpose should be rare.

Corneal topography is considered medically necessary only when the diagnosis of monocular diplopia is thought to be caused by a corneal irregularity.

Corneal topography is covered when the medically necessary criteria above is met, only if the results will assist in defining further treatment. It is not covered for Medicare Advantage Plans and not medically necessary for Commercial Products for routine follow-up testing.

Repeat testing is only indicated if a change of vision is reported in connection with one of the above listed conditions.

Services performed for screening purposes or in the absence of associated signs, symptoms, illness, or injury as indicated above, are not covered for Medicare Advantage Plans and not medically necessary for Commercial Products.

Corneal topography is not covered for Medicare Advantage Plans and not medically necessary for Commercial Products if it is performed pre- or post-operatively in relation to a non-covered procedure (i.e., refractive eye surgery).

Corneal topography is not covered for Medicare Advantage Plans and not medically necessary for Commercial Products for the management of individuals with the following indications (not an all-inclusive list) because computerized corneal topography has not been shown to alter the clinical management of these conditions as the evidence is insufficient to determine the effects of the technology on health outcomes:

- Acanthomoeba keratitis
- Accommodative disorders
- Band keratopathy
- Diplopia
- Epithelial ingrowth following laser in situ keratomileusis (LASIK)
- Interstitial keratitis
- Kerato-conjunctivitis sicca
- Lattice degeneration of retina
- Lens subluxation (e.g., in Marfan syndrome)
- Limbal dermoids
- Microphthalmia
- Neurotrophic keratoconjunctivitis
- Nodular degeneration of the cornea (e.g., Salzmann's corneal degeneration)
- Ocular graft-versus-host disease
- Ocular surface squamous neoplasia
- Open-angle glaucoma
- Post-herpes simplex virus scarring of cornea
- Refractive errors
- Superficial punctate keratopathy

### COVERAGE

Benefits may vary between groups and contracts. Please refer to the appropriate Benefit Booklet, Evidence of Coverage or Subscriber Agreement for applicable not medically necessary/not covered benefits/coverage.

### **BACKGROUND**

Corneal topography describes measurements of the curvature of the cornea. An evaluation of corneal topography is necessary for the accurate diagnosis and follow-up of certain corneal disorders, such as keratoconus, difficult contact lens fits, and pre- and postoperative assessment of the cornea, most commonly after refractive surgery.

Assessing corneal topography is a part of the standard ophthalmologic examination of some patients. Corneal topography can be evaluated and determined in multiple ways. Computer-assisted corneal topography has been used for early identification and quantitative documentation of the progression of keratoconic corneas, and evidence is sufficient to indicate that computer-assisted topographic mapping can detect and monitor disease.

Various techniques and instruments are available to measure corneal topography: keratometer, keratoscope, and computer-assisted photokeratoscopy.

The keratometer (also referred to as an ophthalmometer), the most commonly used instrument, projects an illuminated image onto a central area in the cornea. By measuring the distance between a pair of reflected points in both of the cornea's 2 principal meridians, the keratometer can estimate the radius of curvature of 2 meridians. Limitations of this technique include the fact that the keratometer can only estimate the corneal curvature over a small percentage of its surface and that estimates are based on the frequently incorrect assumption that the cornea is spherical.

The keratoscope is an instrument that reflects a series of concentric circular rings off the anterior corneal surface. Visual inspection of the shape and spacing of the concentric rings provides a qualitative assessment of topography.

A photokeratoscope is a keratoscope equipped with a camera that can provide a permanent record of the corneal topography.

#### **CODING**

## Medicare Advantage Plans and Commercial Products

Non-computer-assisted corneal topography is considered part of the evaluation and management services of general ophthalmologic services (CPT codes 92002–92014), and therefore this service should not be billed separately. There is no separate CPT code for this type of corneal topography. Non-computer-assisted corneal topography should be considered inclusive to Evaluation and Management services.

The following CPT code(s) are covered when the medical criteria above has been met and when filed with one of the covered ICD-10 codes attached, below:

92025 Computerized corneal topography, unilateral or bilateral, with interpretation and report

Covered ICD-10-CM for CPT Code 92025

#### **RELATED POLICIES**

None

#### **PUBLISHED**

Provider Update, February 2025 Provider Update, June 2024 Provider Update, May 2023 Provider Update, July 2022 Provider Update, June 2021

### **REFERENCES**

- 1. Morrow GL, Stein RM. Evaluation of corneal topography: past, present and future trends. Can J Ophthalmol. Aug 1992;27(5): 213-25. PMID 1393805
- 2. Wilson SE, Klyce SD. Advances in the analysis of corneal topography. Surv Ophthalmol. 1991; 35(4): 269-77. PMID2011820
- 3. Martínez-Abad A, Piñero DP, Ruiz-Fortes P, et al. Evaluation of the diagnostic ability of vector parameters characterizingthe corneal astigmatism and regularity in clinical and subclinical keratoconus. Cont Lens Anterior Eye. Apr 2017; 40(2):88-96. PMID 27931882
- 4. Bhatoa NS, Hau S, Ehrlich DP. A comparison of a topography-based rigid gas permeable contact lens design with aconventionally fitted lens in patients with keratoconus. Cont Lens Anterior Eye. Jun 2010; 33(3): 128-35. PMID 20053579
- 5. Weber SL, Ambrósio R, Lipener C, et al. The use of ocular anatomical measurements using a rotating Scheimpflugcamera to assist in the Esclera® scleral contact lens fitting process. Cont Lens Anterior Eye. Apr 2016; 39(2): 148-53.PMID 264749246.
- 6. DeNaeyer G, Sanders DR, Farajian TS. Surface coverage with single vs. multiple gaze surface topography to fit sclerallenses. Cont Lens Anterior Eye. Jun 2017; 40(3): 162-169. PMID 28336224
- 7. Bandlitz S, Bäumer J, Conrad U, et al. Scleral topography analysed by optical coherence tomography. Cont Lens Anterior Eye. Aug 2017; 40(4): 242-247. PMID 28495356

8. Lee H, Chung JL, Kim EK, et al. Univariate and bivariate polar value analysis of corneal astigmatism measurementsobtained with 6 instruments. J Cataract Refract Surg. Sep 2012; 38(9): 1608-15. PMID 22795977 9. de Sanctis U, Donna P, Penna RR, et al. Corneal Astigmatism Measurement by Ray Tracing Versus Anterior Surface-Based Keratometry in Candidates for Toric Intraocular Lens Implantation. Am J Ophthalmol. May 2017; 177: 1-8. PMID28185842

10. Corneal topography. American Academy of Ophthalmology. Ophthalmology. Aug 1999; 106(8): 1628-38. PMID 10442914



# CLICK THE ENVELOPE ICON BELOW TO SUBMIT COMMENTS

This medical policy is made available to you for informational purposes only. It is not a guarantee of payment or a substitute for your medical judgment in the treatment of your patients. Benefits and eligibility are determined by the member's subscriber agreement or member certificate and/or the employer agreement, and those documents will supersede the provisions of this medical policy. For information on member-specific benefits, call the provider call center. If you provide services to a member which are determined to not be medically necessary (or in some cases medically necessary services which are non-covered benefits), you may not charge the member for the services unless you have informed the member and they have agreed in writing in advance to continue with the treatment at their own expense. Please refer to your participation agreement(s) for the applicable provisions. This policy is current at the time of publication; however, medical practices, technology, and knowledge are constantly changing. BCBSRI reserves the right to review and revise this policy for any reason and at any time, with or without notice. Blue Cross & Blue Shield of Rhode Island is an independent licensee of the Blue Cross and Blue Shield Association.