



EFFECTIVE DATE: 03/03/2015
POLICY LAST UPDATED: 03/03/2015

OVERVIEW

Current techniques for diagnosing and monitoring asthma and predicting exacerbations are suboptimal. Two new strategies, evaluation of exhaled nitric oxide (NO) and exhaled breath condensate (EBC) are proposed. These techniques are also potentially useful in the management of other conditions such as chronic obstructive pulmonary disease and chronic cough. There are commercially available devices for measuring NO in expired breath and various laboratory techniques for evaluating components of EBC.

MEDICAL CRITERIA

Not applicable.

PRIOR AUTHORIZATION

Not applicable.

POLICY STATEMENT

BlueCHiP for Medicare and Commercial

Measurement of exhaled nitric oxide and exhaled breath condensate is considered not medically necessary in the diagnosis and management of asthma and other respiratory disorders including, but not limited to, chronic obstructive pulmonary disease and chronic cough because there is insufficient evidence in the published medical literature to demonstrate its efficacy.

COVERAGE

Benefits may vary between groups/contracts. Please refer to the appropriate Evidence of Coverage or Subscriber Agreement for limitations of benefits/coverage when services are not medically necessary.

BACKGROUND

Asthma is characterized by airway inflammation that leads to airway obstruction and hyper-responsiveness, which in turn leads to characteristic clinical symptoms including wheezing, shortness of breath, cough, and chest tightness. Guidelines for the management of persistent asthma stress the importance of long-term suppression of inflammation using steroids, leukotriene inhibitors, or other anti-inflammatory drugs. Existing techniques for monitoring the status of underlying inflammation have focused on bronchoscopy, with lavage and biopsy, or analysis by induced sputum. Given the cumbersome nature of these techniques, the ongoing assessment of asthma focuses not on the status of the underlying chronic inflammation, but rather on regular assessments of respiratory parameters such as FEV-1 (forced expiratory volume in one second) and peak flow. Therefore, there has been interest in non-invasive techniques to assess the underlying pathogenic chronic inflammation as reflected by measurements of inflammatory mediators.

Two proposed strategies are the measurement of exhaled nitric oxide (NO) and the evaluation of exhaled breath condensate (EBC).

Measurement of Exhaled Nitric Oxide

Nitric Oxide is an important endogenous messenger and inflammatory mediator that is widespread in the human body, functioning, for example, to regulate peripheral blood flow, platelet function, immune reactions, and neurotransmission and to mediate inflammation. While the role of NO in asthma pathogenesis is still

under investigation, patients with asthma have been found to have high levels of exhaled NO, which decreases with treatment with corticosteroids. In biologic tissues, NO is unstable, limiting measurement. However, in the gas phase, NO is fairly stable, permitting its measurement in exhaled air. Exhaled NO is typically measured during single breath exhalations. First, the subject inspires NO-free air via a mouthpiece until total lung capacity is achieved, followed immediately by exhalation through the mouthpiece into the measuring device.

According to a 2009 joint statement by the American Thoracic Society and the European Respiratory Society, there is a consensus that the fractional concentration of FeNO is best measured at an exhaled rate of 50 mL per second (FeNO 50 mL/s) maintained within 10% for more than 6 seconds at an oral pressure between 5 and 20 cm H₂O.⁽¹⁾ Results are expressed as the NO concentration in parts per billion (ppb), based on the mean of 2 or 3 values.

Several devices measuring exhaled NO are commercially available in the United States. In 2003, the U.S. Food and Drug Administration (FDA) cleared for marketing the Nitric Oxide Monitoring System (NIOX[®]), with the indication that measurements of the fractional NO concentration in expired breath (FE-NO)] provide the physician with means of evaluating an asthma patient's response to anti-inflammatory therapy, as an adjunct to established clinical and laboratory assessments in asthma. NIOX cannot be used with infants or children approximately under the age of 4, as measurement requires patient cooperation. In March 2008, the NIOX MINO was cleared for marketing. The main differences between this new device and the NIOX are that the NIOX MINO is handheld and portable and that it is not suitable for children younger than age 7 years.

Measurement of Exhaled Breath Condensate

Exhaled breath condensate consists of exhaled air passed through a condensing or cooling apparatus, resulting in an accumulation of fluid. Although EBC is primarily derived from water vapor, it also contains aerosol particles or respiratory fluid droplets, which in turn contain various nonvolatile inflammatory mediators, such as cytokines, leukotrienes, oxidants, antioxidants, and various other markers of oxidative stress. There are a variety of laboratory techniques to measure the components of EBC, including such simple techniques as pH measurement, to the more sophisticated gas chromatography/mass spectrometry or high performance liquid chromatography, depending on the component of interest.

The RTube[™] Exhaled Breath Condensate collection system (Respiratory Research Inc.) and the ECoScreen EBC collection system (CareFusion, Germany) are registered with the FDA as Class I devices that collect expired gas. Respiratory Research has a proprietary gas-standardized pH assay, which, when performed by the company, is considered a laboratory-developed test.

Evidence is insufficient to determine the effect of exhaled nitric oxide and exhaled breath condensate tests on health outcomes, therefore these tests are considered not medically necessary.

CODING

BlueCHiP for Medicare and Commercial

The following CPT codes are considered not medically necessary:

83987

95012

ICD-9 diagnosis codes for which measurement of exhaled nitric oxide and exhaled breath condensate are considered not medically necessary.



Exhaled NO and BC ICD 9 Dx Codes for N

ICD-10 diagnosis codes for which measurement of exhaled nitric oxide and exhaled breath condensate are considered not medically necessary.



Exhaled NO and BC ICD 10 Dx Codes for

RELATED POLICIES

Not applicable.

PUBLISHED

Provider Update, May 2015
Provider Update, June 2014
Provider Update, June 2013
Provider Update, May 2012
Provider Update, May 2011
Provider Update, May 2010
Provider Update, May 2009

REFERENCES

1. Reddel HK, Taylor DR, Bateman ED et al. An official American Thoracic Society/European Respiratory Society Statement: asthma control and exacerbations: standardizing endpoints for clinical asthma trials and clinical practice. *Am J Respir Crit Care Med* 2009; 180(1):59-99.
2. Selby A, Clayton B, Grundy J et al. Are exhaled nitric oxide measurements using the portable NIOX MINO repeatable? *Respir Res* 2010; 11:43.
3. See KC, Christiani DC. Normal values and thresholds for the clinical interpretation of exhaled nitric oxide levels in the US general population: results from the National Health and Nutrition Examination Survey 2007-2010. *Chest* 2013; 143(1):107-16.
4. Schneider A, Schwarzbach J, Faderl B et al. FENO measurement and sputum analysis for diagnosing asthma in clinical practice. *Respir Med* 2013; 107(2):209-16.
5. Katsoulis K, Ganavias L, Michailopoulos P et al. Exhaled nitric oxide as screening tool in subjects with suspected asthma without reversibility. *Int Arch Allergy Immunol* 2013; 162(1):58-64.
6. Sverrild A, Malinowski A, Porsbjerg C et al. Predicting airway hyperreactivity to mannitol using exhaled nitric oxide in an unselected sample of adolescents and young adults. *Respir Med* 2013; 107(1):150-2.
7. Malinowski A, Backer V, Harving H et al. The value of exhaled nitric oxide to identify asthma in smoking patients with asthmalike symptoms. *Respir Med* 2012; 106(6):794-801.

[CLICK THE ENVELOPE ICON BELOW TO SUBMIT COMMENTS](#)

This medical policy is made available to you for informational purposes only. It is not a guarantee of payment or a substitute for your medical judgment in the treatment of your patients. Benefits and eligibility are determined by the member's subscriber agreement or member certificate and/or the employer agreement, and those documents will supersede the provisions of this medical policy. For information on member-specific benefits, call the provider call center. If you provide services to a member which are determined to not be medically necessary (or in some cases medically necessary services which are non-covered benefits), you may not charge the member for the services unless you have informed the member and they have agreed in writing in advance to continue with the treatment at their own expense. Please refer to your participation agreement(s) for the applicable provisions. This policy is current at the time of publication; however, medical practices, technology, and knowledge are constantly changing. BCBSRI reserves the right to review and revise this policy for any reason and at any time, with or without notice. Blue Cross & Blue Shield of Rhode Island is an independent licensee of the Blue Cross and Blue Shield Association.

